

## **Noble Gas Equipment Development – System Technical Specifications and Capabilities**

A high throughput system for processing and detection of radio-xenon for On-Site Inspection (OSI) purposes is currently being developed at FOI. To locate an underground event during an OSI it is important to cover and narrow down a large area of interest in a short time period. This will require a large number of sub-soil gas samples to be analyzed per day. Even if samples are combined, a noble gas system has to have a much higher throughput than currently available. The new system is intended to achieve this and have the capacity of separating high levels of Rn, CO<sub>2</sub> and other gases in combination with the high sensitivity and performances of the current SAUNA II system. The new optimized beta-gamma detector design, and its sensitivity will be presented. Four detectors are collocated in one single lead-shield which is re-designed for simplified field deployment. The improved gas process and its capabilities, e.g. radon separation and re-quantification, will also be covered.

**Primary author:** FRITIOFF, Tomas (Swedish Defence Research Agency, FOI)

**Presenter:** FRITIOFF, Tomas (Swedish Defence Research Agency, FOI)

**Track Classification:** Theme 3: Advances in Sensors, Networks and Processing