

station : Renewal considerations

On the French Polynesian infrasound station I24FR, the wind noise reducing system is based on an 18 m diameter star array with 32 low impedance inlets. In the framework of the station upgrade, the question to use 96 air inlets instead of 32 air inlets is tackled. Global assessment on performances, investment and maintenance will be considered with respect to WNRS.

Primary author: MILLIER, Philippe (CEA - Commissariat à l'énergie atomique et aux énergies alternatives - Département analyse, surveillance, environnement)

Presenter: MILLIER, Philippe (CEA - Commissariat à l'énergie atomique et aux énergies alternatives - Département analyse, surveillance, environnement)

Track Classification: 2. Infrasound Instrumentation